

SERVICE CONSORTIUM, INC. UNIFORMITY DATA SHEET

Sheet No _____ of _____
System No. _____
Location _____

Service Consortium, Inc. • Microlithography Equipment Specialists • 1759 S Main St Su 100 • Milpitas, CA 95035 • (408) 935-9700 • Fax: (408) 719-9700 • www.servcon.com

EXPOSURE _____	APERTURE _____	UNIF TESTER S/N _____	HEAD S/N _____
DATE			
READER			
220 VAC REG			
LAMP VOLTAGE			
LAMP CURRENT			
ETI METER			
PRE-AMP VOLTAGE			
CAL INT @ 0°			
+ 42°			
+ 40°			
+ 35°			
+ 30°			
+ 25°			
+ 20°			
+ 15°			
+ 10°			
+ 5°			
0°			
- 5°			
- 10°			
- 15°			
- 20°			
- 25°			
- 30°			
- 35°			
- 40°			
- 42°			
SUM			
AVG INT $\frac{\text{SUM}}{\text{\# OF READINGS}}$			
HI - LO			
HI + LO			
UNIFORMITY $\frac{\text{HI} - \text{LO}}{\text{HI} + \text{LO}} * 100$			